



NOTE: STRAIN RELIEF MUST NOT SEPARATE UNDER 10 LB RADIAL LOAD.

TOLERANCES UNLESS OTHERWISE SPECIFIED ALL DIMENSIONS IN INCHES DIMENSIONS IN [] METRIC		APPROVALS	DATE
FRACTIONAL $\pm 1/64$		DRAWN T. STUMP	8/11/14
DECIMAL .00 $\pm .01$ [.25mm] .000 $\pm .005$ [.127mm]		CHECKED <i>[Signature]</i>	8/15/14
ANGULAR $\pm .5^\circ$		PROJ ENGR	
		APPROVED <i>[Signature]</i>	8/19/14
		PROJECT	FSCM

PlasticsOne[®]

Medical Design and Manufacture
6591 MERRIMAN ROAD, S.W., ROANOKE, VA, 24018
PHONE: 540-772-7950 FAX: 540-989-7519

TITLE		ANCHOR, A12	
SCALE		4:1	
SIZE	DWG	REV	SHT OF
A	2384-50	A	OF

A	RELEASE PER ECO 22907 TAS 8/7/14	8/21/14	BJR
REV.	DESCRIPTION	APPROVAL DATE	APPROVED BY

THIS DRAWING CONTAINS CONFIDENTIAL AND PROPRIETARY INFORMATION AND MAY NOT BE ISSUED OR REPRODUCED, IN WHOLE, OR IN PART, WITHOUT THE EXPRESS WRITTEN PERMISSION OF PLASTICS ONE INC.